



The Günther Tulip Vena Cava Filter is **MR Conditional**. A patient with this filter can be scanned safely after placement under the following conditions:

- Static magnetic field of 3.0 Tesla or 1.5 Tesla.
- Maximum spatial magnetic gradient of 1600 Gauss/cm (16.0 T/m) or less.
- Normal operating mode.
- Maximum MR system reported, whole-body-averaged specific absorption rate (SAR) of 2.0 W/kg for 15 minutes of scanning or less (i.e., per scanning sequence).

MR image quality may be compromised if the area of interest is within approximately 21 mm of the position of the Günther Tulip Vena Cava Filter.

Additional information is available at www.cookmedical.com

AL-1203-427

Patient name _____

Date of placement _____

Hospital _____

Physician name _____

Physician phone # _____



This patient has
received a Günther
Tulip Vena Cava Filter.
More information
on back side.



Günther Tulip™
VENA CAVA FILTER

Cook Incorporated

P.O. Box 489
Bloomington, IN 47402 U.S.A.
Phone: 800.457.4500

William Cook Europe ApS

Sandet 6, 4632 Bjaeverskov
Denmark
Phone: +45 56 86 86 86

William Cook Australia Pty. Ltd.

Brisbane Technology Park,
12 Electronics Street
Brisbane, QLD 4113 Australia
Phone: +61 7 38 41 11 88