

**ADVIA Chemistry XPT  
Sample Probe Cover for Aptio by Siemens, Aptio by Inpeco and FlexLab  
Automation Interfaces**

---

Our records indicate that your facility may have an ADVIA Chemistry XPT connected to Aptio by Siemens, Aptio by Inpeco or FlexLab Automation system, with an incorrect sample probe cover noted in Table 1.

**Table 1. ADVIA Chemistry XPT Affected Product(s)**

<b>Product</b>	<b>Siemens Material Number (SMN)</b>
HVN Mech Automation Kit - Aptio	11241418

**Reason for Field Action**

Siemens Healthcare Diagnostics is informing customers that when an ADVIA Chemistry XPT instrument, SMN 10723034, is connected to Aptio by Siemens, Aptio by Inpeco or FlexLab Automation system, the required external sample probe cover, SMN 11241418, may need replacing due to a design defect.

**Risk to Health**

There is no risk of user injury or impact on results due to this defect.

**Actions to be Taken by the Customer**

No action is required on your part.

A Siemens Customer Service Engineer will contact you to schedule an inspection of the cover and, if necessary, order and install a re-designed cover.

Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.

Please retain this letter with your laboratory records, and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Customer Care Center or your local Siemens technical support representative.

ADVIA Chemistry XPT is a trademark of Siemens Healthcare Diagnostics.