

DIAsource ImmunoAssays[®] S.A.
Rue du Bosquet 2
1348 Louvain-La-Neuve
Belgium
Tel +32 10 84 99 11
Fax +32 10 84 99 90
www.diasource.be

FSCA Ref: CAPA30680
FSN Ref: CAPA30680

Date: 04.06.2019

Urgent Field Safety Notice

VIP RIA (#RB311/#RB311RUO)

To the attention of the users of DIAsource VIP RIA assay (#RB311/RB311RUO) - lot 191601

Contact details of local representative*

DIAsource ImmunoAssays -Rue du Bosquet 2, B-1348 Louvain-la-Neuve + 32 10 84 99 11



DIASource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA30680
 FSN Ref: CAPA30680

Urgent Field Safety Notice (FSN)

VIP RIA (#RB311/#RB311RUO)

1. Information on Affected Devices*	
1.	1. Device Type(s)* Radioimmunoassay
1.	2. Commercial name(s) VIP RIA
1.	3. Unique Device Identifier(s) (UDI-DI) NA
1.	4. Primary clinical purpose of device(s)* Radioimmunoassay for the in vitro quantitative measurement of vasoactive intestinal polypeptide (VIP) in human plasma. For professional use within a laboratory
1.	5. Device Model/Catalogue/part number(s)* RB311/RB311RUO
1.	6. Software version NA
1.	7. Affected serial or lot number range Lot 191601 (expiry 28/06/2019)
	8. Associated devices NA

2 Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem* Low CPM values measured for the calibrators and kit controls generated with the kit lot 191601, which might lead to an unexpected flat calibration curve and a loss of discrimination between the calibrators. For that reason, the values of the 2 kit controls might fall out of the acceptable ranges defined by DIASource, leading to invalid assays.
2.	2. Hazard giving rise to the FSCA* The results generated with that batch 191601 do not meet the performances expected in the Instructions for Use. Despite this problem, no risk of patient misdiagnosis is being expected in the case the kit controls measured by the laboratories are out of the



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA30680
 FSN Ref: CAPA30680

	acceptable range. As indicated in the Instruction for use, " In order to enable the laboratory to completely monitor the consistent performance of the assay, the following important factors should be checked. 1.The found concentrations of the controls should be within the limits given on the labels of the vials".
2.	3. Probability of problem arising Moderate.
2.	4. Predicted risk to patient/users None
2.	5. Further information to help characterise the problem NA
2.	6. Background on Issue DIAsource registered some customers' complaints related to low CPM values measured for the calibrators and kit controls generated with the kit lot 191601. DIAsource's preliminary investigations highlighted that the tracer lot 19D08 could be responsible for this issue, which leads to a flat calibration curve and a loss of discrimination between the calibrators.
2.	7. Other information relevant to FSCA A follow-up will be provided by the end of the week 23.

	3. Type of Action to mitigate the risk*
3.	<p>1. Action To Be Taken by the User*</p> <p> <input type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input type="checkbox"/> Return Device <input type="checkbox"/> Destroy Device <input checked="" type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Follow patient management recommendations <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU) <input type="checkbox"/> Other <input type="checkbox"/> None </p> <p>1. Stop the use of the un-opened VIP RIA boxes from lot 191601 in customers' inventory until DIAsource provides new tracers to the customers in replacement for on-site device modification;</p>



DIASource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA30680
 FSN Ref: CAPA30680

	2. Discard the tracer 19D08 from the kits lot 191601 and use the remaining reagents only in combination with the new tracer that DIASource will provide 3. Communicate to DIASource the number of un-opened kits lot 191601 in the end-users facilities 4. Communicate to DIASource the number of kits who failed. 5. Complete and send back the Field Safety Notice Customer Reply Form provided on the following pages	
3.	2. By when should the action be completed?	11/06/19
3.	3. Particular considerations for: IVD Is follow-up of patients or review of patients' previous results recommended? No No risk for the patients as far as the kit controls have been measured out of the acceptance criteria. The runs who have achieved correct kit controls values can be validated and the patients samples can be released	
3.	4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes
3.	5. Action Being Taken by the Manufacturer <input type="checkbox"/> Product Removal <input checked="" type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input type="checkbox"/> Other <input type="checkbox"/> None 1) Pursue the tests to find the root cause and to correct the problem 2) Discard the incriminated lot cited above from inventory; 3) Analyze the risks for the patient health. 4) Offer replacements materials to the customers	
3	6. By when should the action be completed?	14/06/19
3.	7. Is the FSN required to be communicated to the patient /lay user?	No
3	8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet? NA NA	



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

FSCA Ref: CAPA30680
 FSN Ref: CAPA30680

4. General Information*	
4.	1. FSN Type* New
4.	2. For updated FSN, reference number and date of previous FSN NA
4.	3. For Updated FSN, key new information as follows: NA
4.	4. Further advice or information already expected in follow-up FSN? * YES
4	5. If follow-up FSN expected, what is the further advice expected to relate to: Delivery timeframe of the replacements products
4	6. Anticipated timescale for follow-up FSN By 07/06/2019
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)
	a. Company Name DIAsource ImmunoAssays
	b. Address Rue du Bosquet 2 , B-1348 Louvain-la-Neuve
	c. Website address www.diasource.be
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *
4.	9. List of attachments/appendices: NA
4.	10. Name/Signature Preud'homme Valérie Technical Support Specialist




DIAsource ImmunoAssays® S.A.
Rue du Bosquet 2
1348 Louvain-La-Neuve
Belgium
Tel +32 10 84 99 11
Fax +32 10 84 99 90
www.diasource.be

FSCA Ref: CAPA30680
FSN Ref: CAPA30680

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback..*</p>

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium


 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

Field Safety Notice Distributor/Importer Reply Form

1. Field Safety Notice (FSN) information	
FSN Reference number*	CAPA30680
FSN Date*	CAPA30680
Product/ Device name*	VIP RIA
Product Code(s)	RB311 & RB311RUO
Batch/Serial Number (s)	191601

2. Distributor/Importer Details	
Company Name*	
Account Number	
Address*	
Shipping address if different to above	
Contact Name*	
Title or Function	
Telephone number*	
Email*	

3. Return acknowledgement to Sender	
Email	Valerie.Preudhomme@diasource.be
Distributor/Importer Helpline	+3210849923
Postal Address	Rue du Bosquet , 2, B-1348 Louvain-la-Neuve, Belgium
Web Portal	https://www.diasource-diagnostics.com/
Fax	+3210849990
Deadline for returning the Distributor/Importer reply form*	11/06/2019



DIAsource ImmunoAssays® S.A.
 Rue du Bosquet 2
 1348 Louvain-La-Neuve
 Belgium
 Tel +32 10 84 99 11
 Fax +32 10 84 99 90
 www.diasource.be

4. Distributors/Importers (Tick all that apply)		
<input type="checkbox"/>	*I confirm the receipt, the reading and understanding of the Field Safety Notice.	Distributor/Importer to complete or enter N/A
<input type="checkbox"/>	I have checked my stock and quarantined inventory – enter number of un-opened kits and date complete.	Distributor/Importer to enter quantity and date
<input type="checkbox"/>	I have identified customers that received or may have received this device	
<input type="checkbox"/>	I have attached customer list	
<input type="checkbox"/>	I have informed the identified customers of this FSN	Date of communication:
<input type="checkbox"/>	I have received confirmation of reply from all identified customers	
<input type="checkbox"/>	I need new Tracers in replacement for on-site modification of the lot 191601	Distributor/Importer to enter quantity
<input type="checkbox"/>	I need new kits in replacement for the kits lot 191601 that failed	Distributor/Importer to enter quantity
Print Name*		
Signature*		
Date *		

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.