

Customer  
Hospital  
City  
Postal code  
Country  
*Attn.: XXX*

## **Urgent Field Safety Notice**

### **AQT90, Risk of patient mix-up**

Dear Customer

#### **Background**

Radiometer has become aware that the barcode readers in the bottom edge of the display and behind the inlet wheel may misinterpret the contents of a locally printed barcode label used for entering patient ID or accession number into the analyzer in connection with a sample measurement.

This relate to barcode types not using a check digit.

A check digit enables the barcode reader to validate the data read, and hence, to capture if the barcode has been misinterpreted and consequently to reject such data.

For barcodes without a check digit, the following factors may add to the risk of misinterpretation:

- Poor paper quality,
- Poor printer quality,
- Improper handling of the barcode (e.g. the barcode is folded or exposed to liquid).

Radiometer uses a barcode type with a check digit when producing barcode labels used for e.g. sampler identification.

#### **Risk for the patient**

Misinterpretation of a barcode, including e.g. patient ID or accession number, may result in the following hazardous situations:

1. Patient mix-up
2. Loss of sample resulting in delayed medical treatment

#### **Affected product**

All AQT90 analyzers.

#### **Your actions**

Radiometer kindly requests you to:

- Check if your institution is using barcode types without a check digit, including e.g. patient ID or accession number, to be read on the AQT90.
  1. If you use barcode types **without** a check digit, Radiometer recommends to either:
    - Enable the check digit on the barcode type currently used, or
    - Change type of barcode to one that includes a check digit,as this enables the barcode reader to validate the data read, and hence, to capture if the barcode has been misinterpreted and consequently to reject such data.

2. If you use barcode types **with** a check digit the barcode reader already validates the data read, and hence, captures if the barcode has been misinterpreted and consequently rejects such data. Hence, no short-term action required.
  3. If you do not use barcodes to be read by the AQT90, no short-term action required.
- Complete the Recall Response Form (last page of this letter) and submit to your Radiometer distributor within two weeks from receipt of this letter.

**Solution provided by Radiometer**

Radiometer will provide the following solution to your analyzer irrespectively of which barcode type is used, or no barcodes are used at all:

For both barcode readers the solution is to make the barcode readers scan the barcode three times (as opposed to twice previously). Further, the data read during the three scans must be identical.

For barcode types without a check digit, this change may potentially reduce the risk of misinterpretation, if caused by the factors, which may add to the risk of misinterpretation, mentioned above.

Implementation of the above change in both barcode readers requires an update of the analyzer software. Radiometer is currently working on the software update and will contact you when released.

**Your help is appreciated**

If you are not the end-user of the affected product, please ensure that this letter is distributed to the final end-user.

If you have any questions, please contact your Radiometer distributor.

Best regards,  
<Radiometer distributor>

# Recall Response Form

Concerning:

## **AQT90, Risk of patient mix-up**

- I have received the customer advisory letter, and can confirm that:
  - We use a barcode type **without** a check digit. Further:
    - We acknowledge that Radiometer recommends to either enabling the check digit on the barcode type currently used, or changing type of barcode to one that includes a check digit, but we cannot currently do this.
    - We will either enable the check digit on the barcode type currently used, or change type of barcode to one that includes a check digit.
  - We use barcode types **with** a check digit. Hence, no short-term actions performed.
  - We do not use barcodes to be read by the AQT90. Hence, no short-term actions performed.

Please note, that Radiometer will provide the solution described to your analyzer irrespectively of the answers above.

Hospital Name:	
Your Name:	
Date:	
Signature:	
Email Address:	