

FSN Ref: FSN\_09%NaCIPFS\_NL\_05Feb2021

FSCA Ref : FSCA\_09%NaCIPFS\_NL\_05Feb2021

Date: 05-02-2021

**Urgent Field Safety Notice**  
**Steriflush® Prefilled 0.9% Sodium Chloride Syringes and Procedure**  
**Packs Containing Steriflush® Prefilled 0.9% Sodium Chloride**  
**Syringes**

For Attention of\*: End User

Contact details of local representative (name, e-mail, telephone, address etc.)\*

Mediq NL

Rijnzahte 10

3454 PV DE MEERN

The Net herlands

+31 653283112

nl.supplyplanning@mediq.com

**Steriflush® Prefilled 0.9% Sodium Chloride Syringes and Procedure Packs Containing Steriflush® Prefilled 0.9% Sodium Chloride Syringes**

**1. Information on Affected Devices\***

1. Device Type(s)*		
<b>Syringe NaCl 0,9% Luer Loek 20 ml *S</b>		
2. Commercial name(s)		
Prefilled syringes 0,9 % sodium chloride		
3. Unique Device Identifier(s) (UDI-DI)		
5608120PFSNACL0.9L-2P2		
4. Primary clinical purpose of device(s)*		
The Sterisets Saline 0.9% <b>NaCl</b> syringes are intended for flushing of intravascular catheters, maintain the patency of indwelling intravascular catheters.		
5. Device Model/Catalogue/part number(s)*		
<b>N/A</b>		
6. Software version		
<b>N/A</b>		
7. Affected serial or lot number range		
<u>Manufacturer's Product Reference Number</u>	<u>Product Name</u>	<u>LOT</u>
14363SAL	Syringe NaCl 0,9% Luer Loek 10 ml	Batch numbers up until 1809XXX (batch number is YYMM123)
14360S	Syringe 3 ml with 0,9% sodium chloride sterile	
14361S	Syringe 5 ml with 0,9% sodium chloride (3ml fill) sterile	
14366S	Syringe 10 ml with 0,9% sodium chloride (5ml fill) sterile	
14362S	Syringe 5 ml with 0,9% sodium chloride	
14363S	Syringe 10 ml with 0,9% sodium chloride	
14364S	Syringe 20 ml with 0,9% sodium chloride	
TP-50-10	Prefilled syringes 0,9% Natriumchloride(0,9%NaCl)	
14363SNSAL	Syringe NaCl 0,9% Luer Loek 10 ml	
14365SNSF	Syringe LL 10 ml NaCl 0,9% (3ml fill) sterile White Cap	
14360SNS	Syringe 3 ml with 0,9% sodium chloride sterile/non sterile	
14361SNS	Syringe 5 ml with 0,9% sodium chloride (3ml fill) sterile/non sterile	
14362SNS	Syringe 5 ml with 0,9% sodium chloride sterile/non sterile	
14363SNS	Syringe 10 ml with 0,9% sodium chloride sterile/non	

		sterile	
	14364SNS	Syringe 20 ml with 0,9% sodium chloride sterile/non sterile	
	14365SNS	Syringe 10 ml with 0,9% sodium chloride (3ml fill) sterile/non sterile	
	14366SNS	Syringe 10 ml with 0,9% sodium chloride (5ml fill) sterile/non sterile	
	100470	Na-und Abschluss-Set / homepump	
	10606912	Dialysis fistel saet	
	10606914	Haemodialyse start-stop saet	
	10606902	Aansluitset Dialyse	
	10606913	Dialyse aansluitset	
	10606901	Afsluitset dialyse	
	10606911	Disconnection set dialysis home care incl. Flushing saline	
	10609006	Oncology set	
	10601030	Port Skiftesaet	
	10609007	Connectionset oncologisc	
	10605810	Picc-line saet	
	8. Associated devices		
	N/A		

<b>2 Reason for Field Safety Corrective Action (FSCA)*</b>	
	1. Description of the product problem*
	Unlikely presence (<0,1%) of trace metals in the syringe stopper used in Steriflush® Prefilled 0.9% Sodium Chloride Syringes, which could potentially generate extremely small brown particles.
	2. Hazard arising from the FSCA*
	<b>N/A</b>
	3. Probability of problem arising
	Possible to occur - however, the risk is being mitigated to Improbable .
	4. Predicted risk to patient/users
	No serious injuries and or death could occur due to the failure mode associated with this.
	5. Further information to help characterise the problem
	<b>N/A</b>
	6. Background on Issue
	Brown particles have been found inside the prefilled syringe containing 0.9%NaCl. After investigation we can conclude that there was an interaction between the sodium chloride 0.9% and the trace metals present in the rubber stopper. The health risk associated with this issue is small as no incident or patient safety has ever been involved.
	7. Other information relevant to FSCA
	<b>N/A</b>

<b>3. Type of Action to mitigate the risk*</b>	
<p><b>1. Action To Be Taken by the User*</b></p> <p> <input type="checkbox"/> Identify Device    <input type="checkbox"/> Quarantine Device    <input type="checkbox"/> Return Device    <input type="checkbox"/> Destroy Device  <input type="checkbox"/> On-site device modification/inspection  <input type="checkbox"/> Follow patient management recommendations  <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU)  <input type="checkbox"/> Other                      <input type="checkbox"/> None </p>	
<p>2. By when should the action be completed? 05-03-2021</p>	<p>Specify where critical! to patient/end user safety</p>
<p>3. Particular considerations for:                      Choose an item.</p> <p>Is follow-up of patients or review of patients' previous results recommended? No</p> <p>Provide further details of patient-level follow-up if required or a justification why none is required</p>	
<p>4. Is customer Reply Required? * (If yes, form attached specifying deadline for return) 05-03-2021</p>	<p>Yes</p>
<p><b>5. Action Being Taken by the Manufacturer</b></p> <p> <input type="checkbox"/> Product Removal                      <input type="checkbox"/> On-site device modification/inspection  <input type="checkbox"/> Software upgrade                      <input type="checkbox"/> IFU or labelling change  <input type="checkbox"/> Other - Provide further details of the action(s) identified. <input type="checkbox"/> None </p> <p>Eg how the manufacturer became aware; brief details of relevant incidents; root cause if known; rationale for containment of problem to only affected devices; other risk mitigation or longer-term preventative action etc.</p>	
<p>6. By when should the action be completed? 05-03-2021</p>	<p>Specify where critical! to patient/end user safety</p>
<p>7. Is the FSN required to be communicated to the patient /lay user?</p>	<p>No</p>
<p>8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet? No</p>	

<b>4. General Information*</b>	
1. FSN Type*	<b>New</b>
2. For updated FSN, reference number and date of previous FSN	N/A
3. For Updated FSN, key new information as follows:	
Please, check the syringe before use. It is necessary to destroy the syringes with presence of brown particles.	
4. Further advice or information already expected in follow-up FSN? *	No
5. If follow-up FSN expected, what is the further advice expected to relate to:	
6. Anticipated timescale for follow-up FSN	FINAL actions completed
7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
Steripack S.A	Only necessary if not evident on letter-head.
Zona Industrial 1, Lote 11 a 14 4560-164 Guilhufe, Penafiel Portugal	Only necessary if not evident on letter-head.
.....	Only necessary if not evident on letter-head.
8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. * yes	
9. List of attachments/appendices:	If extensive consider providing web-link instead.
10. Name/Signature	..... ..... ..... ..... ..... .....

---

**Transmission of this Field Safety Notice**

---

N/A

---

Note: Fields indicated by \* are considered necessary for all FSNs. Others are optional.

**Contact manufacturer**

**Steripack S.A**

Att.: .....  
Zona Industrial 1, Lote 11 a 14  
4560-164 Guilhufe, Penafiel  
Portugal  
Tel.: +351 255 711 355  
Fax: +351 255 711 357  
Web site: [www.sterisets.eu](http://www.sterisets.eu)  
E-mail: .....

**Acknowledgment of receipt**

Sterisets Medica! Products requires an acknowledgment of receipt of this notice.

With regards,

**Steripack S.A**

.....

.....

