

Product Name	Part No.	Lot No.	Quantity
<b>ComposiTCP 30 Interference screw - D 7 - L:25 mm; cannula 1.2m</b>	<b>110004605</b>	<b>190491</b>	<b>3 defects / 147 sent</b>

**Date: September 30, 2021.**

**Attention: To all relevant parties within your organization or, where applicable, to any organization to which the medical device has been transferred and is affected by this measure.**

**Description of the problem:**

**SBM has initiated a voluntary product recall concerning a reference of ComposiTCP Interference screws. According to our traceability system, you have received at least one of the three devices mentioned above. You are therefore concerned by this action.**

**3 cases of: 1 screw Ø9/L35mm Lot 190068 inside a box for screw Ø7/L25mm Lot 190491. Date of manufacture: 2019-01-02; Use before: 2022-01-01.**

Screw Ø9:L35mm

Screw Ø7:L25mm on the box



**Predicted risk to patient/users: Delay in surgery of more than 30mn - Potential impact on surgical technique.**

**Actions to be taken by the user:**

*Include, as appropriate:*

- Identifying and quarantining the medical device,*
- Method of recovery, disposal or modification of device,*
- Recommended patient follow up, e.g. implants,*
- Timelines,*
- Confirmation form to be sent back to the manufacturer if an action is required (e.g. return products).*

**Transmission of this Field Safety Notice:**

This notice needs to be forwarded to all the interested parties within your organization or to any organization to which the device has been transferred (If appropriate).



*SUBJECT*

**Field Safety Notice - PRODUCT  
RECALL**

Reference procedure

**PR 95.12**

*TYPE*

**CAN BE DISCLOSED**

Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

---

**Contact reference person:**

**Régine BAREILLES**

**Coordinatrice Qualité & Affaires Règlementaires**

Quality & Regulatory Affairs Coordinator

**Correspondante Matéiovigilance**

Materiovigilance Correspondent

[regine.bareilles@sbm-fr.com](mailto:regine.bareilles@sbm-fr.com)

**Anne COSPIN**

**Directrice Qualité / Affaires réglementaires**

Quality / Regulatory Affairs Director

[anne.cospin@sbm-fr.com](mailto:anne.cospin@sbm-fr.com)

**S.B.M. SAS**

ZI du Monge

65100 LOURDES

FRANCE

Tél. : (+33) 5 62 42 21 01

Fax : (+33) 5 62 42 21 00

[www.sbm-france.fr](http://www.sbm-france.fr)

**S.B.M. SAS**

ZI du Monge

65100 LOURDES

FRANCE

Tél. : (+33) 5 62 42 21 01

Fax : (+33) 5 62 42 21 00

[www.sbm-france.fr](http://www.sbm-france.fr)