

Helena Biosciences Notification


Electrophoresis Division

CEP026 NOV 2021

Dear Colleague,

Further to customer feedback regarding the V8 Nexus REF: 800008, 800018, and 800020, it has come to our attention that a sub-component of the system REF: 0012.075 Bottom sheet rack/Rack Loading Plate, may have potentially sharp edges.

As such, and as a **precautionary** measure we are asking users to be cautious when handling and cleaning this area whilst our internal investigations are ongoing. As per standard good laboratory practice please ensure appropriate PPE is worn at all times when operating the instrument.



Once our investigations are concluded and if any further corrective or preventative action is required further notifications will be issued.

Email:

Technical Support v8support@helena-biosciences.com

Tel: +44 (0)191 4828440

CEP026

Helena Biosciences Europe
Queensway South,
Team Valley Trading Estate,
Gateshead
NE11 0SD, United Kingdom