

Field Safety Notice

MiniMed™ 600 and 700 series insulin pump

Basal Setting Programming

| | |
|---------------|--|
| MiniMed™ 640G | MMT-1711, MMT-1712, MMT-1751, MMT-1752 |
| MiniMed™ 670G | MMT-1780, MMT-1781, MMT-1782, MMT-1760, MMT-1761, MMT-1762, MMT-1741, MMT-1742 |
| MiniMed™ 720G | MMT-1809, MMT-1810, MMT-1859, MMT-1860 |
| MiniMed™ 740G | MMT-1811, MMT-1812, MMT-1861, MMT-1862 |
| MiniMed™ 770G | MMT-1881, MMT-1882, MMT-1892, MMT-1891 |
| MiniMed™ 780G | MMT-1885, MMT-1886, MMT-1895, MMT-1896 |

January 2022

Medtronic Reference: FA1216

Dear Healthcare Professional,

You are receiving this letter because our records indicate that one or more of your patients have either received a new insulin pump or a replacement insulin pump in the last 6 months. The pump your patient received was NOT pre-programmed with their basal rates or other verified settings (i.e., bolus wizard settings, sensor settings, etc.), which must be set up and saved on their pump prior to use.

Patients receiving a new or replacement insulin pump are receiving a letter that directs them to confirm that their settings have been saved and, if not, to program their settings. The letter reminds them that once the basal rates are entered, they must scroll down to select "**Done**" and then select "**Save**" on the next screen to activate the basal rate settings. Not selecting "**Save**" will result in no basal insulin delivery and can potentially cause severe hyperglycemia which may lead to life-threatening diabetic ketoacidosis (DKA). Your patient(s) may be contacting you to verify what those settings are as shown below in the section titled "**MEDTRONIC PROVIDED THE FOLLOWING INSTRUCTIONS TO PATIENTS**".

Serious injuries have been reported with the use of the MiniMed™ 600 series and MiniMed™ 700 series insulin pumps which may be directly attributed to not setting basal rates. In addition, one death has been reported, although a review by independent clinical experts did not directly attribute this to not setting basal rates. If basal rates are not set in the pump when they should be, it could potentially lead to those events as explained above.

Medtronic

ACTIONS REQUIRED BY YOU:

1. If contacted by your patient, please assist your patients in locating and verifying their prescribed settings on their insulin pumps.
2. Ensure that your patients' verified settings are correctly programmed and saved on their pumps.

MEDTRONIC PROVIDED THE FOLLOWING INSTRUCTIONS TO PATIENTS:

New Users with New Device:

- 1. Do not use your pump until you have consulted with your healthcare professional to determine the settings.**
- 2. Program your settings as described in Steps 4 (c) and (d) below.**

Existing Users: Replacement or Upgrade devices

3. Verify current basal rate settings

To check the current basal rate settings in your pump, follow the instructions on the pump user guide for your pump model.

4. Check if the basal rate settings are present on your pump

If the basal rate settings are present on your pump:

- a. No action is required. For future reference, you may also save your settings to CareLink™, or write them down on a paper and keep it securely.

If the basal rate settings are not present on your pump, please take all the following actions:

- b. Locate the settings for your pump, including basal rate settings, and consult with your healthcare professional to verify they are the most recent settings.
 - i. If you cannot get in touch with your healthcare professional, but your previous settings were uploaded to CareLink™ in the past 90 days, you may log into your CareLink™ Personal, navigate to "Reports", then "Select custom range" to choose a week that had the previous pump's upload, select "DEVICE SETTINGS SNAPSHOT", and select "Generate reports". The settings should have a non-zero basal rate.
- c. Program your new or replaced insulin pump with all your verified settings. Refer to the pump user guide for detailed instructions on programming your insulin pump. If you have your settings but require assistance programming your pump, please contact [<our Helpline / your Medtronic contact at <XXXXX >](#).

Medtronic

- d. As stated in the user guide, during programming the basal settings on your pump, make sure you respond to all pump screens to ensure your basal settings are saved. As shown in the screen sequence below, you **must** first scroll down to select “**Done**”, and then select “**Save**” on the following screen. The settings are successfully saved when the message “**Changes saved**” is shown on the screen.



The exact Basal rates shown above are for example only.

The Competent Authority of your country has been notified of this action.

At Medtronic, patient safety is our top priority, and we are committed to delivering safe and effective therapies. We apologize for any inconvenience this issue may cause you and we appreciate your time and attention in reading this important notification.

Any patients who have further questions or need assistance can contact our Helpline at: <XXXXXXX>.

Sincerely,

Local /BU Manager

Enclosure:

- Pump User Letter