

Rev 1: September 2018

FSN Ref: FSN-2022-004

FSCA Ref: FSN-2022-004

Date: 16 March 2022

Urgent Field Safety Notice (FSN)

**Thermo Scientific™ Remel™ Culti-Loops™ Candida albicans ATCC™ 10231™
PK/5**

For Attention of*: Lab Managers

Contact details of local representative (name, e-mail, telephone, address etc.)* E.mail : mbd.vigilance@thermofisher.com Telephone: +44(0) 1256 841144 Fax: +44(0) 1256 479525

Urgent Field Safety Notice (FSN)

Thermo Scientific™ Remel™ Culti-Loops™ Candida albicans ATCC™ 10231™ PK/5

1. Information on Affected Devices*	
1.	1. Device Type(s)* IVD
1.	2. Commercial name(s) Remel™ Culti-Loops™ Candida albicans ATCC™ 10231™ PK/5
1.	3. Unique Device Identifier(s) (UDI-DI) 00848838038075
1.	4. Primary clinical purpose of device(s)* This product is a ready to use, disposable bacteriological loop containing stabilized viable microorganisms and is recommended for use in the performance testing of culture media, stains, diagnostic kits and reagents, for the maintenance of stock cultures and in the evaluation of bacteriological procedures.
1.	5. Device Model/Catalogue/part number(s)* R4601503
1.	6. Software version N/A
1.	7. Affected serial or lot number range 294388
1.	8. Associated devices N/A

2. Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem* An internal technical investigation has confirmed that Remel™ Culti-Loops™ Candida albicans ATCC™ 10231™ PK/5 R4601503 Lot 294388 may exhibit lower than expected or no growth of <i>C. albicans</i> ATCC 10231.
2.	2. Hazard giving rise to the FSCA* Lower than expected or no growth of <i>C. albicans</i> .
2.	3. Probability of problem arising Low to medium
2.	4. Predicted risk to patient/users There should be no immediate or long-term health consequences from use of this product. The <i>C. albicans</i> culture is a quality control strain. It should have little or no effect on the ability of clinical specimens to grow such a fungus on standard primary fungal isolation media. The clinical risk should be considered negligible.
2.	5. Further information to help characterise the problem N/A
2.	6. Background on Issue An internal investigation has found that R4601503 lot 294388 is not performing as expected.
2.	7. Other information relevant to FSCA Lot. 294388 with the expiry of 11-Nov-2022

4. General Information*		
4.	1. FSN Type*	New
4.	2. For updated FSN, reference number and date of previous FSN	N/A
4.	3. For Updated FSN, key new information as follows:	
	N/A	
4.	4. Further advice or information already expected in follow-up FSN? *	Not planned yet
4	5. If follow-up FSN expected, what is the further advice expected to relate to:	
	N/A	
4	6. Anticipated timescale for follow-up FSN	N/A
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Remel Inc.
	b. Address	12076 Santa Fe Trail Drive Lenexa KS 66215
	c. Website address	www.thermofisher.com
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *	
4.	9. List of attachments/appendices:	Customer Response Form
4.	10. Name	xxx
	Signature	

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p>

Customer Reply Form

1. Field Safety Notice (FSN) information			
FSN Reference number*	2022-004		
FSN Date*	16 March 2022		
Product/ Device name*	Thermo Scientific™ Remel™ Culti-Loops™ Candida albicans ATCC™ 10231™ PK/5		
Product Code(s)	R4601503		
Batch/Serial Number (s)	294388		
2. Customer Details			
Account Number			
Organisation Name*			
Organisation Address*			
Department/Unit			
Shipping address if different to above			
Contact Name*			
Title or Function			
Telephone number*			
Email*			
3. Customer action undertaken on behalf of Healthcare Organisation			
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.		
<input type="checkbox"/>	I performed all actions requested by the FSN.		
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed.		
<input type="checkbox"/>	I have returned affected devices - enter number of devices returned and date complete or N/A	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Comments:	
<input type="checkbox"/>	I have destroyed affected devices – enter number destroyed and date complete.	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Qty	Credit <input type="checkbox"/> Replacement <input type="checkbox"/>
		Comments:	
<input type="checkbox"/>	No affected devices are available for return/ destruction		
<input type="checkbox"/>	Other Action (Define):		
<input type="checkbox"/>	I do not have any affected devices.		
<input type="checkbox"/>	I have a query please contact me (e.g. need for replacement of the product).		
Print Name*			
Signature*			
Date*			

Rev 1: September 2018

FSN Ref: FSN-2022-004

FSCA Ref: FSN-2022-004

4. Return acknowledgement to sender	
Email	MBD.vigilance@thermofisher.com
Telephone Number & Fax	Tel : +44(0) 1256 841144 Fax :+44(0) 1256 479525
Postal Address	
Deadline for returning the reply form*	13 April 2022

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.